

Transducer Protector In Vitro Cytotoxicity Test FINAL REPORT

Client: Finetech Research and Innovation Corporation

Testing Institution: SGS Taiwan Ltd.

Report No.: UB/2013/70737

Report Date: 2013/08/07

The content of this report is invalid if it is not presented as the entire report. Note: 1.

- 2. Any unauthorized alteration, forgery or falsification of the content or appearance of this report is unlawful and offenders may be prosecuted to the fullest extent of the law.
- The results shown in this test report refer only to the test article(s) tested. 3.
- The report is the Chinese version of translations UB/2013/70737A-01 4.

SGS Taiwan Ltd

No. 39, Wu Chyuan 7th Rd., New Taipei Industrial Park, Wu Ku Dist., New Taipei City 24890, Taiwan

t (866-2) 2299-3939 f (866-2) 2298-1338



STUDY SCHEDULE

In Vitro Cytotoxicity Test

Transducer Protector

Report No.: UB/2013/70737

Study Initiation date: 2013/07/25

Experimental starting date: 2013/07/26

Experimental completion date: 2013/08/02

Study completion date: See Study Director's signature date in the report

Name of study Personnel: Jeff Chen



Testing Institution

Name: SGS TAIWAN LTD.

No. 38, Wu Chyuan 7th Rd., New Taipei Industrial Park, Wu Ku Dist., New Taipei Address:

City 24890, Taiwan (R. O. C.)

Client / Sponsor

Finetech Research and Innovation Corporation Name:

No.29, Anle St., Xiushui Township, Changhua County 504, Taiwan (R.O.C.) Address:

Page 3 of 24



TEST ARTICLE INFORMATION

SGS

INFORMATION FOR TEST ARTICLE/ CONTROL ARTICLE

Sponsor Company Name	Finetech research and innovation corporation				
Sponsor Address	No.29, Anle St., Xiushui Township, Changhua County 50	04, Taiwan (R.O.C.)			
Contract study item	☐ Base on the contract ☐ Others				
Name of Test article/ Control article	Transducer Protector				
Batch/Lot number	Base on the specific number on the package: Base on the date on the package: Base on the arrived date Others:				
Specification & Amount	10pcs/pack * 7packs (e.	g.10ml/bottle * 6 bottles)			
Retention amount (Note 2)	The amount of the same lot is sufficient for One test	Two test (for retention)			
External features	External features: liquid powder tablet capsule Other column	Color: translucent white			
Major components & Purity	Major components: Polypropylene meterial housing with membrane	Purity:			
Solvent and solubility	N/A				
Storage condition	⊠Room temperature 4°C Dry Light sensitive Others				
Expiration date(Note 3)	Date: / / (YYYY/MM/DD) or Period: 2 year@month 0 day				
Attachment(Note 4)	☐ Certificate of Analysis ☐ Material Safety Data Shee ☐ Other : ☐ ☒ No attachment (Note4)	et Stability Test Result			
Sterilization	Has been sterilized □YES ☑NO (If Yes, please select the following item) Methods□EOsterilization□Gamma sterilization□Steam sterilization□Other				
Categorization of devices (The column is only for device used)	Contact with intact skin or mucosa (cumulative conta Short-term (no greater than 4 hr) Long-term (exceeding 4 hr) Maximum duration Implanted device				
Specific requirement (Note 5)	N/A				
Note 1. Above all information is d Note 2. If the sponsor doesn't pro- each batch of test article /control a Note 3. If the effective period is le longer than 5 years, the test article Note 4. Determination and docum- other characteristics of the test arti- Note 5. The test article/control ar- return of the kind of test article/co- the sponsor also can fill in the "spe-	wide the retention of test article/control article, the retention of a reserviticle is the responsibility of the Sponsor, and the responsibility of the Sponsor. The responsibility of the Sponsor article will be retained for 5 years only. The retained till the expression of identity, strength, purity, stability, composition, method of specification of identity, strength, purity, stability, composition, method of specification of identity, strength, purity, stability, composition, method of specification of identity, strength, purity, stability, composition, method of specification of identities are the responsibility of the Sponsor. The human intended article, please indicate in the "special requirement". The human intended article, please indicate in the "special requirement". Note treatment method after test if the test article neeticle is the same as the report number.	piry date. If the effective period is outhers, fabrication, derivation or ad of experiment. For retention or take suggests or dose requested by			

版次: 3.1 試驗-對照物質資料表 Information for test article-control article

發行日期:2013.06.14



TABLE OF CONTENTS

Statement of GLP Compliance	6
Quality Assurance Statement.	7
Archiving	8
Abstract	9
Purpose	10
Experimental Design.	11
Results	15
Conclusion	16
Data Management	17
Deviations and Investigations	18
Protocol Amendments.	18
References	19
Tables	20
Figures	23
Test Article Photo.	24



STATEMENT OF GLP COMPLIANCE

All study activities performed by SGS Taiwan were carried out in compliance with the GLP (Good Laboratory Practices) for Nonclinical Laboratory Studies (Department of Health, Taiwan, 2006), current OECD Principles of Good Laboratory Practice (Organization for Economic Cooperation and Development, Paris, ENV/MC/CHEM (98) 17) and U.S. Food and Drug Administration Good Laboratory Practice Regulations, 21 CFR Part 58. (1987). The study was conducted in accordance with the protocol and standard operating procedures and monitored in conformity with the protocol. All laboratory data were accurately recorded and verified. SGS Taiwan made no GLP compliance claim for characterization and verification of the test article identity and properties; this was the responsibility of the sponsor.

Study Director:

SGS Taiwan Ltd.

Date Completed

2013. 08.16

Facility Manager:

Yuanmin Wen / SGS Taiwan Ltd.

Date Completed



QUALITY ASSURANCE STATEMENT

UB/2013/70737 Transducer Protector In Vitro Cytotoxicity Test

This study was audited by Quality Assurance personnel of SGS Life Science Service. The QA inspection report includes review of study plan, result of a study-based audit and results of audit of raw data and study report. The audit report was issued upon the completion of each testing.

QA:

Melissa Lin / SGS Taiwan Ltd.

2013.08.07

Date Completed

Inspection Type	Inspection date	Study phase	Date to facility manager and study director
Study base	2013/07/25	Draft Protocol	2013/07/25
Study base	2013/08/02	Absorbance detection	2013/08/02
Study base	2013/08/07	Raw data & Draft Final report	2013/08/07

Ultra Trace & Industrial Safety Hygiene



ARCHIVING

All the study-related raw data, records, protocol and the final report will be kept in archives room of SGS (TAIWAN) LTD for 5 years. Furthermore, retention of the test articles will be in Sample Storage Room for 5 years. All of the records and test articles are handled according to GLP guideline. Agent authorized by the sponsor can apply for the review according to SGS procedure.

Address:

No. 38, Wu Chyuan 7th Rd., New Taipei Industrial Park, Wu Ku Dist., New Taipei City 24890, Taiwan
(R. O. C.)

Final report	Final Report Copy
Raw Data	In Vitro Cytotoxicity Test - MTT data sheet
	Application Form
Records	Information for test article-control article
Records	GLP Test Article Control From
	and other supplementary record
Protocol	Protocol



ABSTRACT

In vitro cytotoxicity test was performed in this study to evaluate the biological compatibility of "Transducer Protector", which was provided by Finetech Research and Innovation Corporation. Extraction of test article and treatment of mouse lung fibroblast cells (L929 cells) with test article extracts were performed according to ISO10993-12 and ISO10993-5, respectively. Cell viability determined by MTT assay showed that the test article extract had in average <30% inhibitory effects to the viability of cells. Together with qualitative observations of cell morphology and monolayer confluency, these results suggested that the test article extract induced non-cytotoxicity effect in L929 cells.



PURPOSE

According to the nature and duration of the anticipated contact with human tissues when in use medical device should be carefully tested for biocompatibility to avoid potential physiological damage by toxic substances produced or contaminated during manufacturing. In this study, Transducer Protector is subjected to in vitro cytotoxicity test to evaluate toxicity of substances that could be extracted or released from the medical device. Therefore, the test system is Mouse lung fibroblast cells (L929 cells). The original source was from BCRC. Based on recommendations described in ISO10993-5, quantitative determination of cell viability by MTT assay and qualitative observation of cell morphology and growth density are carried out, followed by concluding level of cytotoxicity according to the scoring criteria listed in the document. These results provide practical information for assessing the in vitro cytotoxicity of the medical device.

Ultra Trace & Industrial Safety Hygiene



EXPERIMENTAL DESIGN

Test System

- A. Cell line: Mouse lung fibroblast cells (L929 cells). The original source was from BCRC60091.
- B. Morphology: Fibroblast
- C. Incubation condition Incubate in Minimum essential medium Eagle with 10% horse serum at 37°C in the presence of 5% CO2

Reagents

- A. Trypsin solution (Gibco, Cat No. 25200-056, Lot No.: 1211608)
- B. Horse serum (Gibco, Cat No. 16050-122, Lot No.: 1131917)
- C. L-Glutamine solution (Gibco, Cat No. 25030-081, Lot No.: 1115717)
- D. Penicillin-Streptomycin solution (Gibco, Cat No. 15140-122, Lot No.: 1209968)
- E. Minimum Essential Medium (MEM, Gibco, Cat No. 10370-021, Lot No.: 1237758)
- F. 3-(4,5-Dimethylthiazol-2-yl)-2,5-diphenyltetrazolium bromide (MTT, Sigma, Cat No. M5655,

Lot No.: MKBHK674V)

- G. Dimethyl sulfoxide (DMSO, Sigma, Cat No. D2650, Lot No.: 059K2300)
- H. Sodium pyruvate, 100x (Sigma, Cat No. S8636, Lot No.: 1171639)
- I. 10X Phosphate buffer solution (UniRegion Bio-Tech, Product No. UR-PBS001-5L, Lot No.: PBS001-5A)

Equipments

- A. Orbital Shaker Incubator-3 (HILES, E-600)
- B. Balance-13 (DENVER, TB-214)
- C. Biological safety cabinet-1 (LABCONCO, 3450801)
- D. CO₂ Incubator-1 (ASTEC, SCA-165DS)
- E. Inverted Microscope-2 (OLYMPUS, CKX41 SF)



Preparation of Test Article and Control Article

A. Test Article

The test article was handled under sterile environment and operated with aseptic technique during preparation. MEM complete medium was used as extraction buffer. The test article was extracted with a ratio of 0.2g/ml in MEM complete medium for 24±2 hours at 37±1°C with constant agitation at 150 rpm per criteria described in ISO10993-12. The pH adjustment, filtration and centrifugation were not conducted.

B. Control Articles

- Blank control: MEM complete medium was as blank control,
- Positive control: Polyurethane film ZDEC (Polyurethane film containing Zinc Di-Ethyldithio-Carbamate, RM-A, Hatano Research Institute, Japan) extracted with 0.1g/1mL MEM complete medium was as positive control,
- Negative control: HDPE film (High Density Poly-Ethylene film, RM-C, Hatano Research Institute, Japan) extracted with 0.1g/1mL MEM complete medium was as negative control.
- Incubation Method: Extractions were performed at 37±1°C for 24±2 hours (ISO10993-12) with constant agitation at 150rpm.

In vitro cytotoxicity test-MTT

Cell incubation

a. Preparation of complete MEM cell culture medium

Complete cell culture medium was prepared by mixing 435 mL of MEM, 5 mL of Penicillin- Streptomycin solution, 5 mL of L-Glutamine solution, 5mL of sodium pyruvate and 50 mL of Horse serum. The completed medium was stored at 4°C.

b. Cell culture

Mouse lung fibroblast cells (L929 cells, Food Industry Research and Development Institute, Strain No. BCRC 60091) were used here for cytotoxicity test. The L929 cells

were grown on a 10-cm dish containing 10 mL of complete MEM medium and incubated at 37±1°C in the presence of 5% CO2. Detachment of the cells was performed by washing the cells with 1 X PBS followed by treatment with 1.0 mL/dish of trypsin solution for 3 minutes at 37±1°C. Enzymatic activity of trypsin was terminated by adding complete MEM medium. Then transferred to new 10-cm dish for subculture.

B. In vitro cytotoxicity test

- a. 100 μL of L929 cell suspension (1×104 cells/well) was transferred into each well of a 96-well cell culture plate. The cells were then incubated at 37±1°C for 24±2 hours in a humidified atmosphere containing 5% CO2.
- b. Culture medium was replaced with 100 µL of test article extracts or controls. The cells were then incubated for another 24 hours. Treatments of the cells with the extracts were performed in triplicates.
- c. Morphology and monolayer confluency of cells were observed under microscope and scored in accordance with ISO10993-5. The scoring criteria were summarized in Table 2.
- d. Following evaluation of cell conditions, the culture medium was aspirated from the plates. 50 µL of the MTT solution was then added to each well and the plate was further incubated for 2 hours \pm 10 mins at 37 \pm 1 °C.
- e. MTT solution was replaced with 100 µL of DMSO. The plate was incubated at room temperature for 25±5 minutes and subsequently subjected to a microplate reader equipped with a 570 nm filter for colorimetric measurement (reference 650 nm).
- f. The triplicate results of MTT assay were presented as mean ± standard deviation (S.D.) and were scored in accordance with ISO10993-5 (as in Table 2). If the mean of cytotoxicity was less than 30%, the result will show "<30%".
- g. Scores of cell morphology, confluency, and inhibition of viability were averaged to give final interpretation of cytotoxicity.



h. At end of the testing, all the test material, raw data, results, and reports were properly maintained under the guidance of Good Laboratory Practice (GLP).

6. Quality criteria

- a. Positive control and negative control
- (1) Positive and negative controls should be included in every cytotoxicity test.
- (2) Positive control was Polyurethane film ZDEC; Negative reference material was HDPE film.
- b. Blank
- Measure the absolute value of optical density, OD₅₇₀, The acceptance criteria of blank was ≥ 0.2.
- (2) Blanks were placed both at the left side (row 2) and the right side (row 11) of the 96-well plate.
- (3) The left and right mean of the blanks do not differ by more than 15% from the mean of all blanks.

RESULTS

1. Appearance

The extracts of the test article was not different than the blank control.

2. Cell Morphology

As shown in Table 3 and Figure 1, the cells exposed to negative control showed no significant

change in cell morphology compared to that of reagent control and resulted in a score as 0.

Positive control extract caused severe cellular damage and obvious morphological alteration in

almost all cells. Therefore, the positive control experiment was scored as 4. The cells treated with

test article extract showed discrete intracytoplasmatic granules and no cell lysis. Therefore, the cell

morphology was scored as 0.

3. Monolayer cell confluency

Table 4 and Figure 1 showed that the cells exposed to both reagent control and negative control

extracts have similar growth density. Thus, we scored negative control as 0. Treatment of the cells

with positive control extract abolished cell growth and resulted in approximately 30% of unhealthy

cells remaining. We scored the effect of positive control on cell confluency as 3. In comparison to

reagent control, monolayer confluency of the cells treated with test article extract was determined

as 100% with a score as 0.

Inhibition of cell viability

The acquired readings of OD570 absorbance of reagent control were averaged and set as 0%

inhibition of cell viability. In proportion to reagent control, we determined inhibition of cell

viability of negative control, positive control, and test article as <30%, 96.74% ± 0.52% and <30%,

respectively. The relative values of inhibition of cell viability and their scoring were summarized

in Table 5.



CONCLUSION

The scores of the cytotoxicity test, including morphological evaluation, observation of monolayer confluency, and relative inhibition of cell viability, were averaged and listed in Table 6. Based on the averaged score was concluded that the "Transducer Protector" extract induced non-cytotoxicity to L929 cells according to ISO10993-5.





DATA MANAGEMENT

The quantitative data are showed as mean ± standard deviation (S.D.) and are scored using "Criteria for scores in cytotoxicity test" (Table 2). The qualitative data are scored using "Criteria for scores in cytotoxicity test". The individual score represents the average of triplicates. Mean score is the average of the quantitative and qualitative scores.



DEVIATIONS AND INVESTIGATIONS

There was no deviation and investigation during the test period of this study.

PROTOCOL AMENDMENTS

There was no protocol amendment during the test period of this study.



SGS Taiwan Ltd

No. 36, Wu Chyuan 7th Rd., New Taipei Industrial Park, Wu Ku Dist., New Taipei City 24890, Taiwan

Member of the SGS Group (Société Générale de Surveillance)

t (886-2) 2299-3939 f (886-2) 2298-1338

台灣檢驗科技股份有限公司

Report No.: UB/2013/70737 Date: 2013/08/07



REFERENCES

- 1. Good Laboratory Practice for Nonclinical Laboratory Studies. Title 21 of the U.S. Code of Federal Regulations, Part 58 (1997) United States Food and Drug Administration.
- 2. ISO 10993 (2009) Biological evaluation of medical device Part 5: Tests for in vitro cytotoxicity.
- 3. Mendes SC, Reis RL, Bovell YP, Cunha AM, Blitterswijk CA, Bruijn JD (2001) Biocompatibility testing of novel starch-based materials with potential application in orthopaedic surgery: a preliminary study. Biomaterials. 22, 2057-2064.
- 4. Abiraman S, Varma HK, Kumari TV, Umashankar PR, John A (2002) Preliminary in vitro and in vivo characterizations of a sol-gel derived bioactive glass-ceramic system. Bull. Mater. Sci. 25(5), 419-429.
- 5. ISO 10993 (2012) Biological evaluation of medical devices-Part 12: Sample preparation and reference materials.
- 6. Current OECD Principles of Good Laboratory Practice (Organization for Economic Cooperation and Development, Paris, ENV/MC/CHEM (98) 17).
- 7. EOMP-USL-0027 Operating procedures of the biosafety cabinet and laminar flow, UV-lamp verification and aerobic plate counts. Version 2.2
- EOMP-USL-0030 Maintenance and operating procedures of the microscope. Version 1.0



Table 1 - Summary of extraction ratio for medical device

Thickness (mm)	Extraction ratio (surface area or mass/volume)	Examples of forms of materials Film, sheet, tubing wall	
< 0.5	6 cm ² /mL		
0.5 to 1.0	3 cm ² /mL	Tubing wall, slab, small moulded items	
> 1.0	3 cm ² /mL	Larger moulded items	
> 1.0	1.25 cm ² /mL	Elastomeric closures	
Irregularly shaped solid devices	0.2 g/mL	Powder, pellets, foam, non-absorbent, moulded items	
Irregularly shaped porous devices (low-density materials)	0.1 g/mL	Membranes, textiles	

NOTE: While there are no standardized methods available at present for testing absorbents and hydrocolloids, a suggested protocol is as follows:

Page 20 of 24

Ultra Trace & Industrial Safety Hygiene

⁻determine the volume of extraction vehicle that each 0.1 g or 1.0 cm² of material absorbs;

⁻then, in performing the material extraction, add this additional volume to each 0.1 g or 1.0 cm2 in an extraction mixture.



Table 2 - Scoring criteria for cytotoxicity tests

Grade	Reactivity Cell morphological change		Cell confluency	Inhibition of cell viability (MTT assay)	
0	Non-cytotoxic	Discrete intracytoplasmatic granules and no cell lysis.	90-100%	<30%	
1	Slightly-cytotoxic	Not more than 20 % of the cells are round, loosely attached and without intracytoplasmatic granules, or show changes in morphology; occasional lysed cells are present.		30-40%	
2	Mildly-cytotoxic	Not more than 50 % of the cells are round, devoid of intracytoplasmatic granules, no extensive cell lysis.	40-60%	40-60%	
3	Moderately-cytotoxic	Not more than 70 % of the cell layers contain rounded cells or are lysed; cell layers not completly destroyed.	20-40%	60-80%	
4	Severely-cytotoxic	Nearly complete or complete destruction of the cell layers.	<20%	80-100%	

Page 21 of 24



Table 3 - Scores of cytotoxicity of test article extract in L929 cell morphology

Extracts	Exp 1	Exp 2	Exp 3
Reagent control	0	0	0
Positive control	4	4	4
Negative control	0	0	0
UB/2013/70737	0	0	0

Table 4 - Cytotoxic effect of test article extract in monolayer L929 cell confluency (%)

Extracts	Exp 1	Exp 2	Exp 3
Reagent control	100	100	100
Positive control	30	30	30
Negative control	100	100	100
UB/2013/70737	100	100	100

Table 5 - Cytotoxic effect of test article extract in inhibition of L929 cell viability (%)

Extracts	Exp 1	Exp 2	Exp 3	Mean <u>+</u> SD	Score
Reagent control	<30%	<30%	<30%	<30%	0
Positive control	97.11%	96.15%	96.97%	96.74%± 0.52%	4
Negative control	<30%	<30%	<30%	<30%	0
UB/2013/70737	<30%	<30%	<30%	<30%	0

Table 6 - Summary of cytotoxicity test results

Extracts	Cell morphology	Cell confluency	Inhibition of viability	Mean score	Cytotoxicity
Reagent control	0	0	0	0	None
Positive control	4	3	4	3.67	Severely
Negative control	0	0	0	0	None
UB/2013/70737	0	0	0	0	None

Page 22 of 24

Report No.: UB/2013/70737 Date: 2013/08/07



Cell Morphology




Figure 1 - Morphology and confluency of L929 cells after being exposed to test article or control extracts for 24 hour

Page 23 of 24

SGS

TEST ARTICLE PHOTO



Page 24 of 24